ResMed

S9[™] Series AutoSet[™] / Elite[™] / Escape[™]

POSITIVE AIRWAY PRESSURE DEVICES

Data Management Guide

English

The following table shows where data from the S9 Series devices can be viewed. Data displayed in ResScan[™] can be downloaded via:

- S9 USB adapter and cable, connected directly from the S9 Series device to the computer
- SD card.

Note: Please refer to your S9 Series device Clinical Guide for further details.

Viewing Data

Parameter	S9 Sleep Report Screen	ResScan ¹			
Type of Data	STORED	STORED			
S9 AutoSet / S9 Elite					
AHI/AI/CAI	\checkmark	\checkmark			
OAI/UAI		\checkmark			
Events (sec)		\checkmark			
Flow Limitation (round to flat)		\checkmark			
Flow (L/min)		\checkmark			
Leak (L/sec)	\checkmark	\checkmark			
Minute Ventilation (L/min)		\checkmark			
Pressure (cm H ₂ O)	\checkmark	\checkmark			
Pulse Rate (beats/min) ²		\checkmark			
Snore (quiet to loud)		\checkmark			
Oxygen saturation (SpO_2) (%) ²		\checkmark			
Usage	\checkmark	\checkmark			
S9 Escape					
Usage	\checkmark	\checkmark			

 In ResScan, via the SD card, data for all parameters is available for viewing, as shown in the table above. Via the S9 USB adapter and cable, data available for viewing includes only values for AHI/AI/CAI/OAI/UAI, Leak, Pressure, and Usage.

2. Only available if an oximeter is used with the S9 device.

ResScan Downloaded Data

	Via S9 USB Adapter	Via SD Card
S9 AutoSet / S9 Elite	365 summary sessions	365 summary sessions 30 detailed sessions 7 high-rate detailed sessions
S9 Escape	365 summary sessions	365 summary sessions

ResScan Review Screen Displays

	Statistics	Summary Graphs	Detailed Graphs	Oximetry Statistics ¹
S9 AutoSet / S9 Elite	\checkmark	\checkmark	\checkmark	\checkmark
S9 Escape	\checkmark	\checkmark		

1. Only available if an oximeter is used with the S9 device.

ResScan Review Screen Display Descriptions

	Statistics	Summary Graphs	Detailed Graphs
Apnoea Indices	Shows the AHI, AI, HI, OAI, CAI and UAI for the selected sessions in the Data Browser.	Shows a vertical bar graph where the lower segment is the median Al per hour, and the upper segment is the median AHI per hour.	Shows a cumulative total of the number of apnoeas and hypopneas that have occurred.
Events			Apnoeas are shown at the time they end. The duration of the apnoea, in seconds, is displayed above the symbol. Apnoeas are represented by colored symbols, where the height of the symbol is proportional to the duration of the apnoea. The type of apnoea (obstructive, central, or unknown), is indicated by the symbol and its color. The duration of the apnoea, in seconds, is displayed above the symbol Hypopneas are recorded and displayed after ten seconds.
			Hypopneas are represented by a blue rectangle.
Flow Limitation			Shown on a scale ranging from flat to round.
Flow			Shown as a blue trace.
Leak	Shows the maximum, 95th percentile, and median statistics for the selected sessions in the Data Browser.	Shows the maximum, 95th percentile, and median statistics for single sessions.	Shown as a blue trace. A red line provides a reference level of the recommended maximum acceptable leak.
Minute Ventilation			Shown as a blue trace.
Pressure	Shows the maximum, 95th percentile, and median statistics for the selected sessions in the Data Browser.	Shows the maximum, 95th percentile, and median statistics for single sessions.	Shown as a blue trace.
Pulse Rate			Shown as a blue trace.
Snore			Shown on a scale ranging from quiet to loud.
Oxygen Saturation (SpO ₂)			Shown as a blue trace. A red line provides the 90% reference level to assist identification of desaturations.
Usage	Total hours used, Daily usage, Used Days ≥ X:YY hours, Used Days < X:YY hours, Total days and % Used Days. Calculated for the sessions selected in the Data Browser.	Each period is shown as a solid bar. A hollow bar indicates a period of usage where the end- time is unknown. There is a limit on the maximum number of separate bars shown for a single session.	

Updating Settings

Parameter	Mode		ResScan via	ResScan via
	CPAP	AutoSet	S9 USB Adapter	SD Card
S9 AutoSet	\checkmark	\checkmark		
S9 Elite	\checkmark			
S9 Escape	\checkmark			
Therapy Mode	\checkmark	\checkmark	\checkmark	\checkmark
Set Pressure	\checkmark		\checkmark	\checkmark
Maximum Pressure		\checkmark	\checkmark	\checkmark
Minimum Pressure		\checkmark	\checkmark	\checkmark
Start Pressure	\checkmark	\checkmark	\checkmark	\checkmark
Maximum Ramp Time	\checkmark	\checkmark	\checkmark	\checkmark
Ramp Time	\checkmark	\checkmark	\checkmark	\checkmark
EPR	\checkmark	\checkmark	\checkmark	\checkmark
EPR Level	\checkmark	\checkmark	\checkmark	\checkmark
EPR Inhale	\checkmark		\checkmark	\checkmark
Mask	\checkmark	\checkmark	\checkmark	\checkmark
Tube	\checkmark	\checkmark	\checkmark	\checkmark
Climate Control	\checkmark	\checkmark	\checkmark	\checkmark
Sleep Quality ¹	\checkmark	\checkmark	\checkmark	\checkmark
SmartStart ^{™ 1}	\checkmark	\checkmark	\checkmark	\checkmark
Language	\checkmark	\checkmark	\checkmark	
Local Date & Time	\checkmark	\checkmark	\checkmark	
Reminders	\checkmark	\checkmark	\checkmark	\checkmark

1. Only able to be set on the S9 Elite and S9 AutoSet.

Markov ResScan Detailed Graphs Specifications

Parameter		2	Sampling period (sec)		
	Resolution	Range	via SD Card		
S9 AutoSet / S9 Elite					
Events (sec)	1	10–120 (apnoea duration)	Aperiodic		
АНІ	1	0–120 (events per hour)	Aperiodic		
Flow Limitation	n/a	Round to flat	2		
Flow (L/min)	1	-127 to 127	25Hz		
Leak (L/sec)	0.02	0–5	2		
Minute Ventilation (L/min)	0.1	0–180	2		
Therapy Pressure (cm H ₂ O)	0.2	4–20	2		
Pulse Rate (beats/min) ¹	1	18–300	1		
Snore	n/a	Quiet to loud	2		
SpO ₂ (%) ¹	1	0–100	1		

1. Only available if an oximeter is used with the S9 device.

Glossary

Apnoea

An apnoea is the temporary absence or cessation of breathing. An apnoea is scored when there is reduction in breathing by 75% of the baseline breathing for at least 10 seconds.

ResScan shows three types of apnoeas:

Central Apnoea

A central apnoea is an apnoea during which the upper airway remains open.

Obstructive Apnoea

An obstructive apnoea is an apnoea during which there is a physical closing of the upper airway.

Unknown Apnoea

An unknown apnoea is an apnoea during which a leak higher than 30 L/min occurs, precluding accurate determination of whether the apnoea is obstructive or central.

Apnoea Indices

For all indices, the value shown for Statistics is the total number of events divided by Daily Usage.

 AHI — Apnoea–Hypopnea Index The total number of events is calculated by adding the number of apnoea and hypopnea events.

For graphs, the AHI count is incremented at the occurrence of every event and reset every hour.

- AI Apnoea Index
- HI Hypopnea Index
- CAI Central Apnoea Index
- OAI Obstructive Apnoea Index
- UAI Unknown Apnoea Index

Daily Usage

Daily Usage is total usage in a single session (a session starts at midday and finishes 24 hours later).

• Average Daily Usage

Average daily usage is the result of the sum of Daily Usage divided by Used Days, over a selected time period.

Median Daily Usage

Median Daily Usage is the middle value for daily usage, where values for Daily Usage are listed from low to high, over a selected time period. While a few exceptionally high or low values can have a significant influence on an average measure, the median is typically more reflective of the true central tendency.

Events

An event is the occurrence of a residual apnoea or hypopnea.

Events are recorded as they occur. The maximum number of events stored per session is 500.

Flow

Flow is an estimate of the airflow entering the lungs.

It is derived by taking the total flow and then removing the leak and mask vent flow components.

Flow Limitation

Flow Limitation is a measure of partial upper airway obstruction.

This measure is based on the shape of the inspiratory flow-time curve. A flat shape suggests upper airway obstruction.

Hypopnea

A hypopnea is an episode of shallow breathing during sleep. A hypopnea is scored when there is a reduction in breathing by 50% of baseline breathing with partial upper airway obstruction for 10 seconds or more. The event is scored after 10 seconds of the hypopnea.

Leak

Leak is an estimate of the total rate of air escaping due to mouth and mask leaks.

It is derived by analysing the inspiratory and expiratory airflows, together with the expected mask vent flows.

High or changing leak rates may affect the accuracy of other measurements.

Minute Ventilation

Minute ventilation is the volume of air breathed in (or out) within any 60-second period.

Pulse Rate

The number of heart beats in a 60-second time frame. The pulse rate is calculated by an attached oximeter.

Snore Index

Snore index is a measure based on the amplitude of the pressure wave generated by a patient's snoring.

SpO₂

 ${\rm SpO}_2$ is a measure of the saturation of blood hemoglobin with oxygen, expressed as a percentage. The oxygen saturation is calculated by an attached oximeter.

Therapy Pressure

In CPAP mode, therapy pressure is the set CPAP pressure. In AutoSet mode, therapy pressure is the pressure derived by the AutoSet algorithm.

Tidal Volume

Tidal volume is the volume of air inspired or expired in one respiratory cycle (breath).

Total Hours Used

Total hours used is the total patient usage over a selected time range.

Usage

Usage is the length of time that a patient receives therapy from the device.

The start and end times of the first ten individual periods of usage are available for each session when using ResScan.

Used Days

Used days is the total number of days during which daily usage exceeded the compliance threshold (X hours Y minutes).

% Used Days

% used days calculates the percentage of used days out of the total number of days selected.

Manufacturer: ResMed Ltd 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 Australia Distributed by: ResMed Corp 9001 Spectrum Center Boulevard San Diego CA 92123 USA ResMed (UK) Ltd 96 Milton Park Abingdon Oxfordshire 0X14 4RY UK See www.resmed.com for other ResMed locations worldwide.

ResScan, S9, S9 AutoSet, S9 Elite, S9 Escape and SmartStart are trademarks of ResMed Ltd, and S9, AutoSet and SmartStart are registered in U.S. Patent and Trademark Office. © 2011 ResMed Ltd. 368276/3 2011-03